

DECISION SHEET

Standard(s)- (year and edition): IEC 60601-1:1988 Ed.2 Am1+Am2	Sub clause(s): 15 b)	Sheet n°: DSH-414
Subject: Interference suppression capacitor	Key words: Enclosure, capacitor	Confirmed by CTL at its 39th meeting, in Cologne
Question: In most equipment interference suppression capacitors are not the only capacitors used.		
Decision: Measure the capacitance between and ENCLOSURE and between lines		
Explanatory notes: Replace “interference suppression capacitor” by “the measured capacitance” .		