

DECISION SHEET

Standard(s)- (year and edition): IEC 60601-1:1988 Ed 2 Am1+Am2	Sub clause(s): 6.1, 6.1 n)	Sheet n°: DSH-421
Subject: Marking adjacent to the fuse holder	Key words: Fuse holder, marking	Confirmed by CTL at its 39th meeting, in Cologne
<p>Question:</p> <p>Shall fuse-holders, intended for fuses in accordance with EN60127 (IEC60127), be provided with adjacent complete marking according to the requirements in these standards or can, for example, the rated voltage be omitted? See Clause 6, Marking, of EN 60127-1 (IEC60127-1, First edition, 1988 - Definitions for miniature fuses and general requirements for miniature fuse-links) and EN60127-2 (IEC60127-2, First edition, 1989 - Cartridge fuse-links).</p> <p>Decision:</p> <p>The marking shall be in accordance with the applicable fuse standard.</p> <p>Explanatory notes:</p> <p>This requirement is interpreted as meaning that the marked 'rating' should include the rated current, voltage, fuse characteristic and high (H) or low (L) breaking capacity designation in accordance with the relevant fuse standard.</p>		