



TESTING AND MEASURING EQUIPMENT/ALLOWED SUBCONTRACTING
IEC 60086-1:2011, Edition 11.0
Primary batteries - Part 1: General
IEC 60086-2:2011, Edition 12.0
Primary batteries - Part 2: Physical and electrical specifications

“R” Required
 “S” May be subcontracted, see OD 2012
 “SF” Specialized Facility, see IECCE 02-2
 “W” Witness testing in the categories “MED” and “MEAS”
 “3PPS” Three Phase Power Supply required

Clause	Measurement/testing	Testing / measuring equipment / material needed	Subcontracting
3.19	Small battery	Small battery gauge. Per IEC 60086-1:2011, Figure 1.	R
4.1.3.2	Contact pressure resistance	Steel ball ϕ 1 mm.	R
4.2.4	Open circuit voltage limits	Multimeter. For IEC 60086-1:2011, Table 3.	R
5.3	Conformance check to a specified minimum average duration	Dimension measurement equipment. For IEC 60086-2:2011, Clause 6.	R
		Timer. For IEC 60086-2:2011, Clause 6.	R
		Electric load. For IEC 60086-2:2011, Clause 6.	R
6	Performance - Test condition	Temperature and humidity recorder.	R