



IEC 61010-2-020 Edition (1992) + Amendment 1 (1996)
TESTING AND MEASURING EQUIPMENT/ALLOWED SUBCONTRACTING

R=Required by Lab

S=May be subcontracted

Clause	Measurement/testing	Testing / measuring equipment / material needed	Subcontracting
Annex AA	Dynamic microbiological test method	- Sampling equipment to AA.2.3; Test suspension to AA.2.5 and suitable detection methods	S

Note: The presence of equipment alone does not indicate a satisfactory situation. Assessors must evaluate the equipment design, calibration, uncertainty and documentation to ensure compliance with the directions of the standard. The requirements of ISO Guide 25 regarding validation are applicable, as the tests of this standard are not standardised tests.