



IEC 61010-2-081 - Edition 1
TESTING AND MEASURING EQUIPMENT/ALLOWED SUBCONTRACTING

R=Required by Lab

S=May be subcontracted

Clause	Measurement/testing	Testing / measuring equipment / material needed	Subcontracting
8.2.1 Annex BB	Dynamic mechanical Strength	Impact hammer with an energy of 0.5 J	R

Note: The presence of equipment alone does not indicate a satisfactory situation. Assessors must evaluate the equipment design, calibration, uncertainty and documentation to ensure compliance with the directions of the standard. The requirements of ISO/IEC 17025 regarding validation are applicable, as the tests of this standard are not standardised tests.