



IEC 60601-2-16 - Edition 2
TESTING AND MEASURING EQUIPMENT/ALLOWED SUBCONTRACTING

R=Required by Lab; ; W= Not required in CBTL, can be witnessed in MTL situations
S=May be subcontracted

Clause	Measurement/testing	Testing / measuring equipment / material needed	Subcontracting
19.4		Conductivity meter	W
51.104.2		Blood adjusted to a haematocrit of $0,32 \pm 0,02$	W
51.107		Sound pressure meter laid down in IEC 60651 or IEC 60804 and free-field conditions according to ISO 3744.	W

Note: The presence of equipment alone does not indicate a satisfactory situation. Assessors must evaluate the equipment design, calibration, uncertainty and documentation to ensure compliance with the directions of the standard. The requirements of ISO/IEC 17025 regarding validation are applicable, as the tests of this standard are not standardised tests.