



IEC 61558-2-13, FIRST ED. (1999)
TESTING AND MEASURING EQUIPMENT/ALLOWED SUBCONTRACTING
Safety of power transformers, power supply units and similar devices
Part 2-13: Particular requirements for auto transformers for general use

R=Required by Lab
S=May be subcontracted

Clause	Measurement/testing	Testing / measuring equipment / material needed	Subcontracting
		NOTHING SPECIAL	

Note: The presence of equipment alone does not indicate a satisfactory situation. Assessors must evaluate the equipment design, calibration, uncertainty and documentation to ensure compliance with the directions of the standard. The requirements of ISO Guide 25 regarding validation are applicable, as the tests of this standard are not standardised tests.