



**TESTING AND MEASURING EQUIPMENT/ALLOWED SUBCONTRACTING
IEC 61000-3-3 :2013 Edition 3.0
Electromagnetic compatibility (EMC)**

Part 3-3: Limits – Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current ≤ 16 A per phase and not subject to conditional connection

- “R” Required
- “S” May be subcontracted, see OD 2012
- “SF” Specialized Facility, see IECEE 02-2
- “W” Witness testing in the categories “MED” and “MEAS”
- “3PPS” Three Phase Power Supply required

Clause	Measurement/testing	Testing / measuring equipment / material needed	Subcontracting
6	<i>Measurement of flickers and voltage fluctuations</i>	<i>Power supply, Voltage fluctuation meter, Impedance network, flicker meter</i>	R