



**TESTING AND MEASURING EQUIPMENT/ALLOWED SUBCONTRACTING  
IEC 62722-1, Edition 1 (2014)**

**Luminaire performance – Part 1: General requirements**

- “R” Required
- “S” May be subcontracted, see OD 2012
- “SF” Specialized Facility, see IECEE 02-2
- “W” Witness testing in the categories “MED” and “MEAS”
- “3PPS” Three Phase Power Supply required

Clause	Measurement/testing	Testing / measuring equipment / material needed	Subcontracting
6	Photometric data – LOR	Integrating Sphere or Goniophotometer (with measuring system, calibrated lamps source)	R
	Photometric data – Distribution of luminous intensity	Goniophotometer (with measuring system, calibrated lamps source) Stabilized supply voltage, voltage meter	R R
7	Electrical data	Stabilized supply voltage, watt meter, voltage meter, ampere meter	R
8	Luminaire efficacy data	Nothing Special	--
9	Environmental data	Nothing special	--