



**TESTING AND MEASURING EQUIPMENT/ALLOWED SUBCONTRACTING  
IEC 61000-4-2 : 1995 + A1 : 1998 + A2 : 2000**

- “R” Required
- “S” May be subcontracted, see OD 2012
- “SF” Specialized Facility, see IECEE 02-2
- “W” Witness testing in the categories “MED” and “MEAS”
- “3PPS” Three Phase Power Supply required

| Clause | Measurement/testing                       | Testing / measuring equipment / material needed | Subcontracting |
|--------|---|---|----------------|
| 8.3    | Immunity test to Electrostatic Discharges | ESD simulator, VCP, HCP, reference ground plane | <b>R</b>       |