

Annex 1A Standards of the current accepted scope selected for this Re-assessment – N/A

Annex 1B Initial Assessment / Scope extension Assessment Scope – N/A

Annex 1C Assessment Scope for Specialized Facility

Name of the Test	Standard	Clause of the Standard	Assessment Team acceptance	
			Yes	No
Glow Wire	IEC 60695-2-11	All	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Hot Wire	IECTS 60695-2-20 (See Note)	All	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Ball Pressure	IEC TS60695-10-2	All	<input checked="" type="checkbox"/>	<input type="checkbox"/>
500W Flame	IEC 60695-11-3 IEC 60695-11-20	Method A	<input checked="" type="checkbox"/>	<input type="checkbox"/>
50W Flame	IEC 60695-11-4 IEC 60695-11-10	Method A	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Needle Flame	IEC 60695-11-5	All	<input checked="" type="checkbox"/>	<input type="checkbox"/>
IP	IEC 60529	Codes IP 11 to 68	<input checked="" type="checkbox"/>	<input type="checkbox"/>
			<input type="checkbox"/>	<input type="checkbox"/>